

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product

Switch mode power supplies
(Switching Mode Power Supply for Building-in)

Name and address of the applicant

Astec International Ltd.
16th Floor, Lu Plaza, 2 Wing Yip Street, Kwun Tong
Kowloon, HONG KONG

Name and address of the manufacturer

Astec International Ltd.
16th Floor, Lu Plaza, 2 Wing Yip Street, Kwun Tong, Kowloon, HONG
KONG

Name and address of the factory

Astec Power Philippines Inc.,
Main Avenue Corner Rd "J", Cavite Export Processing Zone, Rosario,,
Cavite, PHILIPPINES

Ratings and principal characteristics

Rated Input : 100-250V, 50/60Hz, 3.5A MAX.
or 120-300VDC, 3A
Rated Output : See test report for details
Protection Class : I
Degree of Protection : IPX0
Remark: The risk management requirements of the standard
were not addressed.

Model/type Ref.

LPQ200A-M, LPQ200C-M, LPQ201-M, LPQ202-M

Additional information (if necessary)

See Test Report for National Differences and Group Differences.

A sample of the product was tested and found
to be in conformity with

IEC 60601-1:2005
IEC 60601-1:2005/AMD1:2012

as shown in the Test Report Ref. No.
which forms part of this certificate

TÜV SÜD PSB Pte Ltd
211-600479-000

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date,

2017-04-11
CBS 17 04 13890 02829



(Yager Bi)



Product Service

TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221