

EU Declaration of Conformance

CE

LumaSense Technologies, Inc.

3301 Leonard Court, Santa Clara, CA, 95054, USA

Declares that the

Model: M310HT BlackBody

Type of Equipment: Blackbody Calibration Source

EN61326:2013, Class A Electromagnetic Emissions

Field Strength 30MHz to 1GHz

Voltage Mains 0.15MHz to 30MHz

EN 61000-3-2 Harmonic current emission

EN 61000-3-3 Fluctuations and flicker

EN61326:2013, Immunity

EN61000-4-2, Electrostatic Discharge Susceptibility

EN61000-4-3, Radiated Electromagnetic Field susceptibility

EN61000-4-4, Power Line Fast Transients Susceptibility

EN61000-4-5, Surge Immunity

EN61000-4-6, Conducted Disturbances Induced by RF Fields

EN 61000-4-11, AC-Power variation and dropout

EN61010-1:2010, Safety requirements for electrical equipment for measurement, control, and laboratory use

2014/35/EU Low Voltage Directive

I, the Undersigned, Hereby Declare That the Equipment Specified Above Conforms to the Directives and Standards as Specified

Connell Yuk

Senior Quality Assurance Manager

LumaSense Technologies

Date August 01, 2018