

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

Component Power Supply

Name and address of the applicant

SL POWER ELECTRONICS CORP
6050 KING DRIVE, BLDG A
VENTURA, CA 93003 USA

Name and address of the manufacturer

SL POWER ELECTRONICS CORP
6050 KING DRIVE, BLDG A
VENTURA, CA 93003 USA

Name and address of the factory

Note: When more than one factory, please report on page 2

BRIDGEPOWER VINA CO LTD
LOT B9 THUY VAN INDUSTRIAL ZONE VIET TRI CITY PHU
THO PROVINCE
VIETNAM Additional Information on page 2

Ratings and principal characteristics

See Page 2

Trademark / Brand (if any)



Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

GB10S09P02, GB10S10P02, GB10SXXP01, KP010SXXP
See Page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to EN 60601-1:2006/ A1:2013/ A12:2014;
National Differences specified in the CB Test Report. Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E116994-D1029-1/A0/C0-CB issued on 2019-11-27

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-12-18

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-35012-UL

Model Details:

GB10S09P02, GB10S10P02, GB10SXXP01, KP010SXXP, where XX representing output 5V to 24V dc.

Factories:

WENDENG JEIL ELECTRONICS CO LTD
XIAMEN ROAD NO.2 WENDENG ECONOMIC DEVELOPMENT ZONE WEIHAI CITY SHANDONG PROVINCE
CHINA

Ratings:

GB10SXXP01, GB10S09P02, GB10S10P02, KP010SXXP
100-240 Vac, 50-60Hz, 0.3A (0.3A - 0.15A)

Rated Output:

GB10 Series: 5Vdc, 2.0A to 24Vdc, 0.41A; KP010SXXP: 5Vdc, 2.0A to 24Vdc, 0.41A
(See model nomenclature in the Test Report, Model Differences section for detail)

Additional Information:

The Risk Management requirements of the standard were not addressed.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-12-18

Signature:

Jolanta M. Wroblewska