

Certificate



Certificate no.

CU 72171574 01

License Holder:

Excelsys Technologies Ltd.
27 Eastgate Drive
Eastgate Business Park
Little Island, Cork
Ireland

Manufacturing Plant:

Shenzhen WATT Electronics Co., Ltd.
No. 5, Tunnel 1, Tangfang Garden
35th District, Baoan County
518101 Shenzhen, Guangdong
China

Test report no.: USA-JE 31792008 001

Client Reference: Diarmuid Hogan

Tested to: ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1
CAN/CSA-C22.2 No. 60601-1:14

Certified Product: Switched Mode Power Supply

License Fee - Units

Model Designation:

7

CX06M-0000-defgh (CoolX CoolPac without CoolMod);
d = N; C; S; P or X; e = '-'; 1, 4 or 5, A-Z; f = A or B;
g = Not used; '-' or L (Not used = Standard model
(if h is not used), '-' = Standard model (if h is used),
L = Cover); h = Optional.
Any alphanumeric character. Logistic use only

contd.

7

Appendix: 1, 1-25

Licensed Test mark:



Date of Issue
(day/mo/yr)
17/07/2017

Certificate



Certificate no.

CU 72171574 02

License Holder:

Excelsys Technologies Ltd.
27 Eastgate Drive
Eastgate Business Park
Little Island, Cork
Ireland

Manufacturing Plant:

Shenzhen WATT Electronics Co., Ltd.
No. 5, Tunnel 1, Tangfang Garden
35th District, Baoan County
518101 Shenzhen, Guangdong
China

Test report no.: USA-JE 31792008 001

Client Reference: Diarmuid Hogan

Tested to: ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1
CAN/CSA-C22.2 No. 60601-1:14

Certified Product: Switched Mode Power Supply

License Fee - Units

contd.

Model designation:

CX06M-wxyz-defgh (CoolX CoolPac with CoolMod);
w, x, y, z = 0, #, A, B, C, D, E or G; d = N; C; S; P or
X; e = '-'; 1, 4 or 5, A-Z; f = A or B; g = Not used;
'-' or L (Not used = Standard model (if h is not used),
'-' = Standard model (if h is used), L = Cover);
h = Optional. Any alphanumeric character.
Logistic use only.

contd.

Licensed Test mark:



Date of Issue

(day/mo/yr)

17/07/2017

Certificate



Certificate no.

CU 72171574 03

License Holder:

Excelsys Technologies Ltd.
27 Eastgate Drive
Eastgate Business Park
Little Island, Cork
Ireland

Manufacturing Plant:

Shenzhen WATT Electronics Co., Ltd.
No. 5, Tunnel 1, Tangfang Garden
35th District, Baoan County
518101 Shenzhen, Guangdong
China

Test report no.: USA-JE 31792008 001

Client Reference: Diarmuid Hogan

Tested to: ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1
CAN/CSA-C22.2 No. 60601-1:14

Certified Product: Switched Mode Power Supply

License Fee - Units

contd.

Model designation:

CX06M-0DC0-X-A-A (complete configuration)

Cma-bcd (CoolX CoolMod)

a = A, B, C, D, E, G; - = Not Used or '-', P, C or S;

bcd = Any three alphanumeric characters.

Optional (not used on standard models). Logistics use only. Note: Use '-' to designate standard model

when bcd is used. e.g. CmB-X03

Not Used when bcd is not used. e.g. CmB

contd.

Licensed Test mark:



Date of Issue

(day/mo/yr)

17/07/2017

Certificate



Certificate no.

CU 72171574 04

License Holder:

Excelsys Technologies Ltd.
27 Eastgate Drive
Eastgate Business Park
Little Island, Cork
Ireland

Manufacturing Plant:

Shenzhen WATT Electronics Co., Ltd.
No. 5, Tunnel 1, Tangfang Garden
35th District, Baoan County
518101 Shenzhen, Guangdong
China

Test report no.: USA-JE 31792008 001

Client Reference: Diarmuid Hogan

Tested to: ANSI/AAMI ES60601-1:2005+A2 (R2012) +A1
CAN/CSA-C22.2 No. 60601-1:14

Certified Product: Switched Mode Power Supply

License Fee - Units

contd.

Rated Voltage: AC 100-240V, 50-60Hz
Rated Current: 5-2.5A
Protection Class: Class I, no applied parts
Output Ratings: 600W max. (at any configuration)

Special Remarks: Replaces certificate CU72161235.01

Licensed Test mark:



Date of Issue

(day/mo/yr)
17/07/2017