

## Declaration of Conformity

Manufacturer: **LumaSense Technologies A/S**  
**Energivej 30**  
**DK-2750 Ballerup, Denmark**

Hereby declares that the products

### **Photoacoustic Multi Gas Monitor – INNOVA 3433i**

From serial no. 140-001

Conforms to the provisions of the EMC Directive 2004/108/EC, Low Voltage Directive 2006/95/EC and RoHS2 Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

And is in compliance with the following standards

EMC: **EN 61326-1: (2013)**: Electrical equipment for measurement, control and laboratory use. EMC requirements – Part 1: General requirements.

Safety: **IEC 61010-1 3<sup>rd</sup> ed. (2010)**: Safety requirements for electrical equipment for Measurement, control and laboratory use.

**CAN/CSA – C22.2 No. 61010-1-12 Third edition**: Safety requirements for Electrical Equipment for Measurement, Control and Laboratory use – Part 1: General requirements.

**UL61010-1 3<sup>rd</sup> ed. (2012)**: Safety requirements for Electrical Equipment for Measurement, Control and Laboratory use – Part 1: General requirements.

This product meets its published specifications at the time of shipment from the factory and has been manufactured in compliance with the provisions of the relevant Lumasense production code and our Quality System. The product has been tested using calibrated equipment that is traceable to national and international standards. The final test has been recorded and is available in the form of test records for examination upon request.

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, pursuant to EN 17050-1. It does not affect the warranty obligations of the supplier.

Place of issue: Ballerup, Denmark

Date of issue: October 2018

Signature: \_\_\_\_\_



Inge Lise B. Olsen, Q.A. Manager