



Ref. Certif. No.

US-161070-M1-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Power Supply
Name and address of the applicant	SL POWER ELECTRONICS CORP 1595 Wynkoop Street Suite 800 Denver, CO 80202 United States
Name and address of the manufacturer	SL POWER ELECTRONICS CORP 1595 Wynkoop Street Suite 800 Denver, CO 80202 United States
Name and address of the factory	INDUSTRIAS S L S A DE C V CIRCUITO SIGLO XXI 2055, COL PARQUE INDUSTRIAL EX-XXI MEXICALI, Baja California 21254 Mexico <input type="checkbox"/> Additional Information on page 2
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	Input rating: 100-240Vac, 50-60Hz, 8.0A <input checked="" type="checkbox"/> Additional Information on page 2
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	NCF660SXXYZZ, Where XX represents the output voltage which may be any number from 12 to 48. Y can be K (for Class I); ZZ can be any number between 00-99, blank or any letter from AA to ZZ, only for market purpose, not influence safety function.
Additional information (if necessary may also be reported on page 2)	The report was revised to include technical modifications. National Differences: CA, US <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	E116994-D1043-1/A1/C0-CB issued on 2026-02-27

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/nbnames

Date: 2026-03-02
Original Issue Date: 2024-12-17

Signature: Mauricio Avila



Ref. Certif. No.

US-161070-M1-UL

Additional Ratings:

For Forced Air Cooled 300LFM Min.:
Main Output: 12Vdc/46.25A Max to 48Vdc/13.65A Max
Standby Output: 5Vdc/1A Max.
For Convection:
Main Output 12Vdc/30.42A Max to 48Vdc/9.1A Max
Standby Output: 5Vdc/1A Max.
Standard Models Output Rating:
NCF660S12K:
For 300LFM: 12V / 46.25A, 5V / 1A
For Convection: 12V / 30.42A, 5V / 1A
NCF660S24K:
For 300LFM: 24V / 27.3A, 5V / 1A
For Convection: 24V / 18.13A, 5V / 1A
NCF660S48K:
For 300LFM: 48V / 13.65A, 5V / 1A
For Convection: 48V / 9.1A, 5V / 1A

Additional Information:

The risk management requirements of the standard were not addressed

Summary of Modifications:

Addition of a new construction.
Update trademark.
Update ratings information.
Update Test Item Particulars section, see CB Test Report for details.

Additional information (if necessary)



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2026-03-02
Original Issue Date: 2024-12-17

Signature:

Mauricio Avila