

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching power supply unit
(Switching Mode Power Supply for Building-in)

Name and address of the applicant

Astec International Limited - Philippine Branch
3rd & 4th Floor, Techno Plaza One Bldg, #18 Orchard Road, Eastwood
City Cyberpark, Bagumbayan
1110 Quezon City, PHILIPPINES

Name and address of the manufacturer

Astec International Limited - Philippine Branch
3rd & 4th Floor, Techno Plaza One Bldg, #18 Orchard Road, Eastwood
City Cyberpark, Bagumbayan, 1110 Quezon City, PHILIPPINES

Name and address of the factory

Astec Power Philippines Inc.
104 Laguna Boulevard, Laguna Technopark, 4026 Sta. Rosa Laguna,
PHILIPPINES
For further information please see attachment

Ratings and principal characteristics

Rated Input : 100-240VAC, 8.5A MAX, 50/60Hz
Rated DC Output : See test report for details
Construction : Built-in
Protection Class : I
Degree of Protection : IPX0

Model/type Ref.

LCM600Q, LCM600W, LCM600Q-T-401, LCM600L,
LCM600U, LCM600W-T-401, LCM600Q-N, LCM600N,
LCM600L-N, LCM600W-N

Additional information (if necessary)

Certificate SG-MD-00450M1 issued on 2015-02-04 is replaced by this version due to technical changes.

A sample of the product was tested and found to be in conformity with


IEC 60601-1:2005
IEC 60601-1:2005/AMD1:2012

as shown in the Test Report Ref. No. which forms part of this certificate

TÜV SÜD PSB Pte Ltd
211-400622-200

This CB Test Certificate is issued by the National Certification Body

Date, 2017-09-25
CBS 17 09 51485 01489


(Yager Bi)



Additional factory information:

Name and address of the factory
Nom et adresse de l'usine

Astec Power Philippines Inc.
Main Avenue Corner Rd "J", Cavite Export Processing
Zone, Rosario, Cavite, Philippines

Zhongshan General Carton Box Factory Co., Ltd.
No. 62, Qi Guan Road West, Shiqi District
528400 Zhongshan City, Guangdong
PEOPLE'S REPUBLIC OF CHINA

Astec Electronics (Luoding) Co. Limited
No. 68 Baocheng Road East, Fucheng Luoding,
527200 Guangdong, PEOPLE'S REPUBLIC OF CHINA

Additional Information:

*The risk management requirements of the
standard were not addressed.*



Date, 2017-09-25
CBS 17 09 13890 01489

Signature : (Yager Bi)



Product Service