

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Switching Power Supply

Name and address of the applicant
Nom et adresse du demandeur

SL POWER ELECTRONICS CORP
BLDG A, 6050 KING DR
VENTURA CA 93003, USA

Name and address of the manufacturer
Nom et adresse du fabricant

SL POWER ELECTRONICS CORP
BLDG A, 6050 KING DR
VENTURA CA 93003, USA

Name and address of the factory
Nom et adresse de l'usine

INDUSTRIAS S L S A DE C V
CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI 21254 MEXICALI BC
MEXICO

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Additional Information on page 2
See Page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

SMT

Model / Type Ref.
Ref. De type

MINT1150VXXYYZWW
See Page 2

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{ème} page

Additionally evaluated to EN 60601-1:2006; National Differences specified in the CB Test Report.

Additional Information on page 2

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

E116994-A92-CB-1 issued on 2012-10-11

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2012-10-11

Signature:

Jolanta M. Wroblewska

For full legal entity names see www.ul.com/ncbnames

Model Details:

MINT1150VXXYYZWW where V is A (Class I construction) or B (Class II construction), XX represents output voltage which may be any number from 12 to 56, YY represents any number 00 through 99, Z represents A to Z, WW represents additional configurations indicating non-safety related options which may be any number 00 through 99.

Factories:

SL XIANGHE POWER ELECTRONICS CORP
NO 4 SHUANGXING NORTH RD XIANGHE ECONOMIC & TECHNOLOGICAL DEVELOPMENT ZONE XIANGHE
HEBEI 065402
CHINA

Ratings:

Input: 100-240 Vac, 50-60 Hz, 2.0 A

Output:

With 200 LFM (linear feet per minute) forced air flow:
Main Output: 12Vdc/12.5A to 56Vdc/2.68A, Maximum 150W

Without forced air flow:

Main Output: 12Vdc/8.33A to 56Vdc/1.79A, Maximum 100W*

*Maximum output power of models MINT1150V12YYZWW is 95W when the input voltage is from 100Vac to 105Vac.

Additional Information:

The product has been evaluated with the exception of Risk Management assessment.


Additional information (if necessary)**Information complémentaire (si nécessaire)**

- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2012-10-11

Signature:



Jolanta M. Wroblewska