



Ref. Certif. No.

US-39485-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switched Mode Power Supply

Name and address of the applicant

EXCELSYS TECHNOLOGIES LTD
UNIT 27
EASTGATE BUSINESS PARK LITTLE ISLAND CO CORK
Ireland

Name and address of the manufacturer

EXCELSYS TECHNOLOGIES LTD
UNIT 27
EASTGATE BUSINESS PARK LITTLE ISLAND CO CORK
Ireland

Name and address of the factory

Shenzhen Watt Electronics Co Ltd
BAOAN BRANCH, 5 TUNNEL 1
TANGFANG GARDEN 35 DISTRICT
BAO'AN SHENZHEN, GUANGDONG, 518101
China

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

Input: 100-240V, 50-60Hz, 15A-7.5A, Class I, no applied part.
Model CX18M - Output: 1800W maximum (refer to test report-specific ratings)

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

CTF Stage 2

Model / Type Ref.

CX18M-000000-defgh, CX18M-uvwxyz-defgh

Additional information (if necessary may also be reported on page 2)

National Differences specified in the CB Test Report.

The risk management requirements of the standard were not addressed

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

E230761-D1006-1/A0/C0-CB issued on 2022-01-20

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-01-31

Signature:

Jolanta M. Wroblewska